






Probe Selection Guide

Model	AM1612	AM1640	AM1660	AM1710	AM1730	AM1751	AM1760	AM1762	AM1210
Image									
Type	Full Immersion PRT	Precision Industrial PRT		Secondary Reference PRT			Secondary SPRT		Reference Type S TC
Temperature Range (°C)	-200 to 160	-200 to 420	-200 to 670	-40 to 160	-200 to 420	-200 to 670	-200 to 670		0 to 1300
Nominal Resistance at 0°C	100 Ω							25 Ω	N/A
Temperature Coefficient	0.00385 Ω / Ω / °C			0.003925 Ω / Ω / °C					N/A
Accuracy	<0.05 °C at 0 °C	<0.035 °C at 0 °C		< 0.012 °C at 0 °C			< 0.006 °C at 0 °C		See data sheet
Long Term Drift*	< 0.04 °C			< 0.01 °C			< 0.004 °C		<0.5 °C at 1210 °C after 1 year typical usage
Short Term Stability	< 0.02 °C	< 0.01 °C		< 0.007 °C			< 0.002 °C		< 0.2 °C at 1084.62 °C
Thermal Shock**	< 0.02 °C	< 0.007 °C		< 0.005 °C			< 0.002 °C		N/A
Hysteresis	< 0.01 °C			< 0.005 °C			< 0.001 °C		N/A
Sheath Material	Stainless Steel	Inconel™		Stainless Steel	Inconel™			Alumina	
Sheath Dimensions (OD x L)	3 mm x 50 mm	0.25 in x 12 in or 0.187 in x 9 in	0.25 in x 12 in	0.25 in x 12 in or 0.187 in x 9 in	0.25 in x 12 in	0.25 in x 12 in or 0.25 in x 20 in	0.25 in x 20 in	0.25 in x 20 in or 0.25 in x 12 in	
Options	N/A	N/A		90° Bend			90° Bend		Cold Junction
Calibration	ISO 17025 accredited calibration included								Test report
Typical Applications	Climate/humidity chambers and freezers for validation and calibration	A robust precision probe for temperature measurement in a variety of media		An affordable secondary reference grade probes designed for use in the laboratory, but also for demanding field calibration and measurement when tighter uncertainties are required			A secondary level SPRT for customers needing a reliable laboratory grade reference probe with tight uncertainty capabilities and long term stability		High temperature thermocouple calibration work, normally reserved for the laboratory, and used as reference TC in high temperature drywell and horizontal thermocouple calibration furnaces.

*For PRTs/SPRTs measured at TPW after 100 hours at max temperature

**For PRTs/SPRTs after 10 thermal cycles from minimum to maximum temperatures

Note:

- 1) Probe selection guide is for reference only, please see probes datasheets for more details
- 2) Carrying cases included for all AM17XX probes (excluding -BEND models)
AM16XX and AM1210 models do NOT include carrying cases

Your Note

Dotted lines for note-taking.